

Medical & Surgical

Traceability of medical implants

Plastic parts are largely used for medical implants, and traceability of those parts is a must. The plastic adopted is the UHMWPE. The wavelength of the Green-Lase radiation, 532nm, is the optimal choice to grant permanent and contaminant-free marking.

VL – GREEN 10W

LENS: 160 mm

MARKING TIME: 31 s

MATERIAL: Plastic, UHMWPE

